
CLEANTYPE® PRIME SERIES

TOUCH+ (WITH BACKLIGHT)

TOUCH







SPECIAL FEATURES

- IP68 rated medical keyboard
- Adjustable brightness of backlight (TOUCH+)
- Fully washable and disinfectable
- 110 keys incl. num pad and track pad
- Waterproof silicone USB protection cap
- 15 cm USB cable with waterproof silicone cap plus 175 cm USB extension cable
- Excellent tactile feeling
- Integrated Clean Mode for cleaning
- Magnetic mounting
- EN60601-1-2 Ed 4 certified

SPECIFICATION

Number of keys:	110 (US-Layout)
Keyswitch technology:	Carbon contact
Switching cycles:	Approx.2 mill. (per key)
Protection level:	IP68
Housing material:	Silicone
Housing color:	White & black
Backlight:	Blue color, 5-level adjustable brightness

Interfaces:	USB 2.0
Cable length:	15 cm + 175 cm extension
Operating temperature:	0°C ~ +55°C
Storage temperature:	-30°C ~ +70°C
Dimensions:	365 (L) x 134 (W) x 14.2 (T) mm
Weight:	Approx.705g
Warranty:	2 years
Certification:	EN60601-1-2 Ed 4, CE, FCC, RoHS, REACH

ORDER INFORMATION

Cat. No.	Product Description	Backlight	Color	Layout
KSI-U10210	CLEANTYPE® PRIME TOUCH+	√	White	US/DE/JP/KR/THAI
KSI-U10211	CLEANTYPE® PRIME TOUCH		White	US/DE/JP/KR/THAI
KSI-U20210	CLEANTYPE® PRIME TOUCH+	√	Black	US/DE/JP/KR/THAI
KSI-U20211	CLEANTYPE® PRIME TOUCH		Black	US/DE/JP/KR/THAI

Other layouts, configurations and interfaces on request. Customized key function on request (Firmware update via USB)

Version 10
2023

GETT Gerätetechnik GmbH
Mittlerer Ring 1
08233 Treuen
Germany
Tel: +49 37468 660 0
E-Mail: info@gett-group.com
Internet: www.gett-group.com

GETT Asia Ltd.
Room 806-7, Kowloon Plaza
No. 485 Castle Peak Road
Kowloon, Hong Kong
Tel: +852 3461 9789
E-Mail: info@gett-asia.com
Internet: www.gett-asia.com

GETT Assembly (Asia) Co., Ltd.
4F & 5F, Building 1, No.97, Fuyong Area 107 Road
Baishixia Community, Fuyong Street
Baoan District, Shenzhen, China
Tel: +86 755 2323 6358
E-Mail: info@gett-asia.com
Internet: www.gett-asia.com

gett-asia.com
gett-group.com

EN60601-1-2 ED 4 CERTIFIED

EN 60601-1-2 Ed 4 applies to the basic safety and essential performance of Medical Equipment (ME) equipment and ME systems in the presence of electromagnetic disturbances and to electromagnetic disturbances emitted by ME equipment and ME systems.

SHORT OPERATING INSTRUCTIONS



Press and hold lock key for 5s to enable Clean Mode. Clean Mode is indicated by breathing backlight.



Adjust backlight level with this button. Five level can be adjusted including off.

CLEANING INSTRUCTION

For further details please refer to below link:
https://gett-group.com/en/gett-service/cleaning-and-material-resistances

GETT Gerätetechnik GmbH
Mittlerer Ring 1
08233 Treuen
Germany
Tel: +49 37468 660 0
E-Mail: info@gett-group.com
Internet: www.gett-group.com

GETT Asia Ltd.
Room 806-7, Kowloon Plaza
No. 485 Castle Peak Road
Kowloon, Hong Kong
Tel: +852 3461 9789
E-Mail: info@gett-asia.com
Internet: www.gett-asia.com

GETT Assembly (Asia) Co., Ltd.
4F & 5F, Building 1, No.97, Fuyong Area 107 Road
Baishixia Community, Fuyong Street
Baoan District, Shenzhen, China
Tel: +86 755 2323 6358
E-Mail: info@gett-asia.com
Internet: www.gett-asia.com